



California Facility Purified Water Specification Testing Result

For the month of: July 2016 (16183-16213)

Listed below are the testing results for the purified water used in the manufacture of products at Hardy Diagnostics.

| Test | Testing Frequency | Units | In-House Specification (method detection limit) | Testing Data |
|---|-----------------------|---|---|--|
| Minimum Resistivity ^{*1,3,4,5} | Continuous Monitoring | Megohm* cm | > 18.0 | 18.21 |
| pH ^{**} | Daily | N/A | 5.5 – 7.5 | 7.30 |
| Total Organic Carbon ^{1,3,5} | Monthly | ug/L | <500 | ND |
| Heavy Metals (Single) ^{1,3,4} (Cd, Cr, Cu, Ni, Pb and Zn) | Annually | mg/L | < 0.05 | Cd – ND Cr – ND Cu – ND Pb – ND Ni –ND Zn – 0.00077 |
| Heavy Metals (Total) ^{3,4} | Annually | mg/L | < 0.1 | 0.00085 |
| Ammonia/OrganicNitrogen ³ | Monthly | mg/L | < 0.1 | ND |
| Total Chlorine Residual ^{3,4} | Monthly | mg/L | <0.1 | ND |
| Maximum Bacterial Content ^{†***1,3,4,5} | Weekly | colony forming units (CFU) per milliliter | <10 | <1.0 |
| Water Quality ^{3,4} ratio | Annually | ratio | 0.8 – 3.0 | 1.11 |
| Use Test (Student <i>t</i>) ³ | Quarterly | N/A | ≤ 2.78 | ≤ 2.78 |
| Inhibitory Residue ⁴ | Annually | N/A | < 15% | 3.2% |
| Maximum Silicate, SiO ₂ ^{3,4} | Annually | mg/L | ≤ 0.05 | ND |

ND = Not Detected at or above the method detection limit.

*Testing data is given as a monthly average at the water source.

**Testing data is given as a monthly average.

***Testing data is given as an average.

References:

1. *Preparation and Testing of Reagent Water in the Clinical Laboratory*, C3-A4. Clinical Laboratory Standards (CLSI), Villanova, PA.
2. *Quality Assurance for Commercially Prepared Microbiological Culture Media*, M22-A3. Clinical Laboratory Standards Institute (CLSI – formerly NCCLS), Villanova, PA.
3. American Public Health Association, *Standard Methods for the Examination of Water and Wastewater*, Washington, D.C.
4. *Manual for the Certification of Laboratories Analyzing Drinking Water*, Criteria and Procedures Quality Assurance, Environmental Protection Agency (EPA).
5. USP. *USP-NF, Water for Pharmaceutical Purposes <1231>*. Rockville, MD: US Pharmacopeial Convention.

Wendy Hadley
Manager of Quality Control

Hardy Diagnostics
1430 W. McCoy Lane
Santa Maria, CA 93455
(800) 266-2222
www.hardydiagnostics.com