



Mast Group Ltd.
Mast House, Derby Road, Bootle
Liverpool, Merseyside, L20 1EA
United Kingdom
Tel: + 44 (0) 151 472 1444
Fax: + 44 (0) 151 944 1332
email: sales@mastgrp.com
Web: www.mastgrp.com

Mast Diagnostica GmbH
Feldstrasse 20
DE-23858 Reinfeld
Germany
Tel: + 49 (0) 4533 2007 0
Fax: + 49 (0) 4533 2007 68
email: mast@mast-diagnostica.de
Web: www.mastgrp.com

Mast Diagnostic
12 rue Jean-Jacques Mention
CS91106, 80011 Amiens, CEDEX 1
France
Tél: + 33 (0) 3 22 80 80 67
Fax: + 33 (0) 3 22 80 99 22
email: info@mast-diagnostic.fr
Web: www.mastgrp.com



MAST ASSURE™ HAEMOPHILUS INFLUENZAE CAPSULE-TYPING ANTISERA

Liquid stable antisera for the determination of capsular-type specific antigens of *Haemophilus influenzae*.

FOR IN VITRO DIAGNOSTIC USE ONLY

Contents: See pack label.

Formulation

MAST ASSURE™ ANTISERA are prepared from rabbits hyperimmunised with standard strains of killed organisms possessing known serotypes or group specific antigens and contain 0.085% sodium azide as preservative.

Stability and storage

Store unopened at 2 to 8°C until the expiry date shown on the pack label. Once opened, MAST ASSURE™ ANTISERA should be stored at 2 to 8°C and may be used until the expiry date given on the label.

Do not freeze reagents.

Warnings and precautions

For *in vitro* diagnostic use only. Observe approved biohazard precautions and aseptic techniques. To be used only by adequately trained and qualified laboratory personnel. Sterilise all biohazard waste before disposal. Sodium azide preservative may be toxic if ingested and may react with lead and copper plumbing to form highly explosive salts. Always dispose of by flushing to drain with plenty of water. Refer to Product Safety Data sheet.

Materials required but not provided

Standard microbiological supplies and equipment such as loops, applicator sticks, clean glass microscope slides or glass test tubes swabs, MAST culture media, incinerators and incubators, etc., as well as reagents and additives such as sterile 0.85% saline solution.

Procedure

Slide agglutination of live organisms

1. Dispense two 5 to 10 µl volumes of sterile 0.85% saline solution (saline) onto a carefully cleaned microscope slide. The slide may be partitioned using a chinagraph pencil. With a platinum wire or disposable inoculation loop take one 1 to 2mm colony of live organisms from a fresh culture on Chocolate Agar and emulsify into each drop of saline to produce a distinct and uniform turbidity.
2. Place a drop (30 to 40µl) of antiserum onto one of the emulsified isolates and on to the other a drop (30 to 40µl) of saline as a control.
Note: Do not allow the organism to contaminate the antiserum dropper bottle.
3. Mix the reagents by tilting the slide back and forth for 60 seconds while viewing it under indirect light against a dark background.

4. Distinct clumping or agglutination within this period, without clumping in the saline control (auto-agglutination), should be regarded as a positive result.

Interpretation of results

Isolates producing a distinct positive reaction with an antiserum are assumed to be a *Haemophilus influenzae* exhibiting the capsule-type specified by antiserum producing the reaction.

If the result is misleading with slight or slow (more than 60 seconds) agglutination, the test should be repeated by selecting another colony where organisms have a larger amount of capsular antigen present.

Limitations of use

Only cultures of organisms identified as *Haemophilus influenzae* by morphological and biochemical features should be serotyped with this product.

Monovalent antisera are intended for use in rapid slide agglutination tests.

Quality control

It is recommended that quality control should be performed with at least one organism to demonstrate a positive reaction and at least one organism to demonstrate a negative reaction. Do not use the product if the reactions with the control organisms are incorrect. Check for signs of deterioration. Do not use reagents if they are contaminated or cloudy.

References

Bibliography available on request.