

HOSPITAL LOGO



Innovating Together™

SNOTBUSTER™ aka SLSolution - HOW TO GUIDE



Remove one SnotBuster™ tube from the plastic package.



Peel open the sterile Sputum Dipper™ pouch and remove the device.



③ Insert the Sputum Dipper™ into the container of sputum specimen.



Completely submerge and rotate the drill-shaped head of the Sputum Dipper™ in the sample for approximately 10 seconds ensuring the plastic coils of the dipper are filled up with sputum sample. If necessary angle the Sputum Dipper™ and tilt the sputum container to ensure the dipper is filled with sample.



Remove the cap from the SnotBuster™ tube and place the Sputum Dipper™ into the tube.



6 Break the shaft at the premolded breakpoint.



Re-cap the SnotBuster[™] tube.



8 Leave at room temperature for 15 minutes.

Note: Prolonged time (up to six hours) will not inhibit bacteria and fungi survival in the sample.



Ovortex the SnotBuster™ tube for 30 seconds at 2000-2500 RPM or until the mixture has liquefied.



Unscrew the cap of the SnotBuster™ tube and aseptically remove an aliquot using a sterile loop. Plant onto appropriate culture medium for the isolation of bacteria and fungi.

Copan Diagnostics Inc.

26055 Jefferson Ave. Murrieta, California 92562 USA

Toll Free: (800) 216-4016 (US & Canada)

Phone: (951) 696-6957 Fax: (951)600-1832

E-Mail: info@copanusa.com

www.copanusa.com