

Instructions for Use

NEUTRALIZING BUFFERED PEPTONE WATER

Cat. no. K281	Neutralizing Buffered Peptone Water (nBPW), 20x125mm Tube, 10mL	20 tubes/box
Cat. no. U181	Neutralizing Buffered Peptone Water (nBPW), 500ml Polycarbonate Bottle, 400mL	10 bottles/box

INTENDED USE

Hardy Diagnostics Neutralizing Buffered Peptone Water (nBPW) is recommended for use in the recovery of sub-lethally injured *Salmonella* species from industrial samples prior to selective enrichment and isolation.

This product is not intended to be used for the diagnosis of human disease.

SUMMARY

Salmonella spp. may be present in foods, particularly poultry products, yet cells may be sub-lethally injured by food processing techniques. Consequently, it may be difficult to recover injured cells of this organism using selective media and the organism may go undetected using traditional culture techniques. Beginning July 1, 2016, the United States Department of Agriculture, Food Safety and Inspection Service (USDA FSIS) instituted new guidelines for the development of Neutralizing Buffered Peptone Water (nBPW) to aid in the recovery of *Salmonella* spp. from domestic and imported poultry verification sampling, including chicken carcass rinses, poultry parts rinses and young turkey carcass sponge swabs.⁽⁶⁾ The USDA FSIS also states nBPW is safe as a direct rinse or swab, and should be used as a non-selective pre-enrichment medium to promote the recovery of sub-lethally injured bacteria, particularly *Salmonella* spp.⁽⁶⁾

nBPW contains peptones that act as nitrogenous compounds to promote bacterial growth. Phosphate salts in the buffer help to maintain pH. Maintenance of pH is important when attempting to recover sub-lethally injured cells, because a low pH can be detrimental to the repair and growth of damaged microorganisms. In addition, nBPW contains neutralizing agents to reduce the inhibitory effects of carryover from microbial interventions and disinfectants.

FORMULA

Ingredients per liter of deionized water:*

Buffered Peptone Water (BPW)	20.0g
Sodium Bicarbonate	12.5g
Lecithin	7.0g
Sodium Thiosulfate	1.0g

Final pH 7.7 +/- 0.5 at 25°C.

* Adjusted and/or supplemented as required to meet performance criteria.

STORAGE AND SHELF LIFE

Storage: Upon receipt, store away from direct light at 2-8°C. away from direct light. Media should not be used if there are any signs of deterioration, discoloration, contamination, or if the expiration date has passed. Product is light and temperature sensitive; protect from light, excessive heat, moisture, and freezing.

The expiration date on the product label applies to the product in its intact packaging when stored as directed. The product may be used and tested up to the expiration date on the product label and incubated for the recommended incubation times as stated below.

Refer to the document "[Storage](#)" for more information.

PRECAUTIONS

This product may contain components of animal origin. Certified knowledge of the origin and/or sanitary state of the animals does not guarantee the absence of transmissible pathogenic agents. Therefore, it is recommended that these products be treated as potentially infectious, and handle observing the usual universal blood precautions. Do not ingest, inhale, or allow to come into contact with skin.

This product is for laboratory use only. It is to be used only by adequately trained and qualified laboratory personnel. Observe approved biohazard precautions and aseptic techniques. All laboratory specimens should be considered infectious and handled according to "standard precautions." Refer to the document "[Guidelines for Isolation Precautions](#)" from the Centers for Disease Control and Prevention.

For additional information regarding specific precautions for the prevention of the transmission of all infectious agents from laboratory instruments and materials, and for recommendations for the management of exposure to infectious disease, refer to CLSI document M29: *Protection of Laboratory Workers from Occupationally Acquired Infections*.

Sterilize all biohazard waste before disposal.

Refer to the document "[Precautions When Using Media](#)" for more information.

PROCEDURE

Sample Collection: Consult reference methods for complete procedures on sample collection.⁽¹⁻⁶⁾

Method of Use: Gently mix nBPW prior to use. Consult listed references for complete procedures for handling poultry carcasses prior to rinsing and swabbing and for information on the recovery of *Salmonella* spp. from food or poultry samples.^(1,2,5,6)

1. Inoculate 10g or 10mL of sample for every 50mL of Neutralizing Buffered Peptone Water (nBPW).
2. Incubate at 35°C. for 18 to 24 hours.
3. Transfer 10mL of the incubated sample to 100mL of Tetrathionate Broth ([Cat. no. U165](#)) and incubate at 35°C. Other selective enrichments may be used.^(1,2,5,6)
4. After 24 and 48 hours, subculture to Brilliant Green Agar ([Cat. no. G75](#)), XLD Agar ([Cat. no. G65](#)) and/or HE Agar ([Cat. no. G63](#)) and incubate plates for 18 to 24 hours at 35°C. NOTE: It is recommended that more than one selective agar be used in parallel, since no single medium is appropriate in all situations, to ensure recovery when salmonellae are present.^(1,2,5,6)
5. Examine plates for typical colonies of *Salmonella* spp. and perform further testing for complete identification. See the IFU for the plated media used.

INTERPRETATION OF RESULTS

Consult listed references for appropriate interpretation of results.⁽¹⁻⁶⁾

Following incubation, examine solid media for growth and typical colony morphology.

LIMITATIONS

It is recommended that biochemical, immunological, molecular, or mass spectrometry testing be performed on colonies from pure culture for complete identification of bacteria and/or fungi.

nBPW is a non-selective medium. Overgrowth of competing flora in the test sample may affect recovery of salmonellae.

Refer to the document "[Limitations of Procedures and Warranty](#)" for more information.

MATERIALS REQUIRED BUT NOT PROVIDED

Standard microbiological supplies and equipment such as loops, other culture media such as Brilliant Green Agar ([Cat. no. G75](#)), XLD Agar ([Cat. no. G65](#)), HE Agar ([Cat. no. G63](#)), or Tetrathionate Broth ([Cat. no. U165](#)), swabs, pipets, applicator sticks, incinerators, and incubators, etc., as well as serological and biochemical reagents, are not provided.

QUALITY CONTROL

Hardy Diagnostics tests each lot of commercially manufactured media using appropriate quality control microorganisms and quality specifications as outlined on the Certificate of Analysis (CofA) and the CLSI document M22-A3 *Quality Assurance for Commercially Prepared Microbiological Culture Media*. The following microorganisms are routinely used for testing at Hardy Diagnostics:

Test Organisms	Inoculation Method*	Incubation			Results
		Time	Temperature	Atmosphere	
<i>Salmonella enterica</i> ATCC® 14028	A	18-24hr	35°C	Aerobic	Growth and typical colony morphology upon subculture to XLD Agar (red colonies with black centers)
<i>Escherichia coli</i> ATCC® 25922	A	18-24hr	35°C	Aerobic	Partial to complete inhibition upon subculture to XLD Agar

* Refer to the document "[Inoculation Procedures for Media QC](#)" for more information.

USER QUALITY CONTROL

End users of commercially prepared culture media should perform QC testing in accordance with applicable government regulatory agencies, and in compliance with accreditation requirements. Hardy Diagnostics recommends end users check for signs of contamination and deterioration and, if dictated by laboratory quality control procedures or regulation, perform quality control testing to demonstrate growth or a positive reaction and to demonstrate inhibition or a negative reaction, if applicable. Hardy Diagnostics quality control testing is documented on the certificate of analysis (CofA) available from Hardy Diagnostics [Certificate of Analysis](#) website. Also refer to the document "[Finished Product Quality Control Procedures](#)," and the CLSI document M22-A3 *Quality Assurance for Commercially Prepared Microbiological Culture Media* for more information on the appropriate QC procedures. See the references below.

PHYSICAL APPEARANCE

Neutralizing Buffered Peptone Water (nBPW) should appear opaque, cloudy, and light yellow in color.

REFERENCES

1. Juven, B.J., N. Cox, J.S. Bailey, J.E. Thomson, O.W. Charles, and J.V. Schutze. 1984. Recovery of *Salmonella* from artificially contaminated poultry feeds in non-selective and selective broth media. *Jour. of Food Prot.* ; 47:299-302.
2. Sadovski, A.Y. 1977. *J. Food Technology*; 12:85-91.
3. American Public Health Association. *Standard Methods for the Examination of Dairy Products*, APHA, Washington, D.C.
4. APHA Technical Committee on Microbiological Methods for Foods. *Compendium of Methods for the Microbiological Examination of Foods*, APHA, Washington, D.C.
5. U.S. Food and Drug Administration. *Bacteriological Analytical Manual*. Arlington, VA
<http://www.fda.gov/Food/FoodScienceResearch/LaboratoryMethods/ucm2006949.htm>

ATCC is a registered trademark of the American Type Culture Collection.

IFU-10860[D]



1430 West McCoy Lane, Santa Maria, CA 93455, USA

Phone: (805) 346-2766 ext. 5658

Fax: (805) 346-2760

Website: HardyDiagnostics.com

Email: TechnicalServices@HardyDiagnostics.com

[Ordering Information](#)

Distribution Centers:

California · Washington · Utah · Arizona · Texas · Ohio · New York · Florida · North Carolina

The Hardy Diagnostics manufacturing facility and quality management system is certified to ISO 13485.

Copyright© 2020 by Hardy Diagnostics. All rights reserved.

HDQA 2207D [D]