

Purified (Reagent) Water Specification Testing Results (Ohio Facility)

For the month of: May 2019

Listed below are the testing results for the purified water used in the manufacture of products at Hardy Diagnostics.

Test	Testing Frequency	Units	In-House Specification (method detection limit)	Testing Data
Minimum Resistivity* 1,3,4,5	Continuous Monitoring	Megohm cm	> 16.0	17.83
pH**	Daily	N/A	5.5 – 7.5	6.01
Total Organic Carbon 1,3,5	Monthly	μg/L	<500	ND
Heavy Metals (Single) 1,3,4 (Cd, Cr, Cu, Ni, Pb and Zn)	Annually	mg/L	< 0.05	Cd – ND Cr – ND Cu – ND Pb – ND Ni – ND Zn – ND
Heavy Metals (Total) 3,4	Annually	mg/L	< 0.1	ND
Ammonia/Organic Nitrogen ³	Monthly	mg/L	< 0.1	ND
Total Chlorine Residual 3,4	Monthly	mg/L	<0.1	ND
Maximum Bacterial Content***	Weekly	colony forming units (CFU) per milliliter	<10	<1
Water Quality 3,4 ratio	Annually	ratio	0.8 - 3.0	0.96
Use Test (Student t) 3	Quarterly	N/A	<u><</u> 2.78	≤2.78
Inhibitory Residue 4	Annually	N/A	< 15%	10.5%
Maximum Silicate, SiO ₂ 3,4	Annually	mg/L	≤ 0.05	ND

ND = Not Detected at or above the method detection limit.

References:

- 1. Preparation and Testing of Reagent Water in the Clinical Laboratory, C3-A4. Clinical Laboratory Standards (CLSI), Villanova, PA.
- 2. Quality Assurance for Commercially Prepared Microbiological Culture Media, M22-A3. Clinical Laboratory Standards Institute (CLSI formerly NCCLS), Villanova, PA.
- 3. American Public Health Association, Standard Methods for the Examination of Water and Wastewater, Washington, D.C.
- 4. Manual for the Certification of Laboratories Analyzing Drinking Water, Criteria and Procedures Quality Assurance, Environmental Protection Agency (EPA).
- 5. USP. USP-NF, Water for Pharmaceutical Purposes <1231>. Rockville, MD: US Pharmacopeial Convention.

Andre Hsiung

Director of Technical Services

Hardy Diagnostics 1430 W. McCoy Lane Santa Maria, CA 93455 (800) 266-2222 www.hardydiagnostics.com

^{*}Testing data is given as a monthly average at the water source.

^{**}Testing data is given as a monthly average.

^{***}Testing data is given as an average.