



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

Hardy Diagnostics
1430 W. McCoy Lane
Santa Maria, CA 93455

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

AC-2546

Certificate Number


ANAB Approval

Certificate Valid Through: 01/16/2022
Version No. 002 Issued: 11/06/2019



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Hardy Diagnostics
 1430 W. McCoy Lane
 Santa Maria, CA 93455
 Alistair Wood 805-361-2695
wooda@hardydiagnostics.com
www.hardydiagnostics.com

CALIBRATION

Valid to: **January 26, 2022**

Certificate Number: **AC-2546**

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Air Samplers (Velocity) ¹	(100, 200) L/min	4 L/min	Wind Tunnel, Reference Anemometer, Barometer, Hygrometer

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2546.



Vice President

