

# Evaluation of the NG-Test CARBA 5 for the Phenotypic Detection of Carbapenemases

Meghan A. Wallace, Albert Liu, Kaitlin F. Mitchell, Carey-Ann D. Burnham

Department of Pathology and Immunology, Washington University School of Medicine, St. Louis, MO

meghan.wallace@wustl.edu  
cburnham@wustl.edu

## INTRODUCTION

As carbapenem-resistant *Enterobacterales* continue to proliferate in the U.S. and throughout the world, rapid and reliable detection of the carbapenemases they produce is a pressing need. Hardy Diagnostics NG-Test CARBA 5 is a rapid, simple, and visual multiplex lateral flow immunochromatographic assay for the detection of the five common carbapenemases (KPC, OXA-48-like, VIM, IMP, and NDM). The test utilizes labeled monoclonal antibodies against the 5 carbapenemases to detect the presence of these enzymes in an organism after incubation in a lysis buffer. Our objective was to evaluate the performance of the Hardy NG-Test CARBA 5 for sensitivity and specificity when tested using isolates cultured on various media types and includes *Enterobacteriaceae* and *Pseudomonas* species.

## MATERIALS AND METHODS

### Isolates.

50 Isolates From the Washington University strain bank

- Barnes Jewish Hospital (n=34)
- Urine and hospital environmental strains from specimens from Pakistan (n=16)

60 Isolates from Hardy Diagnostics

### Test Conditions.

16-24 hours subculture with an ertapenem disk between the 3<sup>rd</sup> and 4<sup>th</sup> quadrant

- TSA with 5% blood (BAP)
- MacConkey Agar (MAC)

CLSI guidelines for antimicrobial susceptibility testing with ertapenem

- Mueller Hinton Agar (MH)

Carbapenemase inactivation method was performed on all isolates

**Quality Control.** Quality control (QC) with the listed organisms was performed on every day of testing. The same processing used for test isolates was followed for QC organisms.


| QC Strain                                  | Expected Result          |
|--|--------------------------|
| <i>Klebsiella pneumoniae</i> ATCC BAA-1705 | Positive for KPC         |
| <i>Klebsiella pneumoniae</i> NCTC 13442    | Positive for OXA-48-like |
| <i>Klebsiella pneumoniae</i> NCTC 13439    | Positive for VIM         |
| <i>Escherichia coli</i> NCTC 13476         | Positive for IMP         |
| <i>Escherichia coli</i> ATCC BAA-2146      | Positive for NDM         |
| <i>Klebsiella pneumoniae</i> ATCC BAA-1706 | Negative                 |

**Comparator Methods.** Each isolate had a genotypic resistance mechanism previously determined using molecular methods (PCR, Gene Xpert Carba-R, or whole genome sequencing (WGS)).

## MATERIALS AND METHODS

Figure 1. NG-Test CARBA 5 Processing

Touch 3 colonies (BAP and MAC) or spots (MH) around the ertapenem disk



## RESULTS

Table 1. Resistance Mechanism Detection by NG-Test CARBA 5 on BAP

| Resistance Mechanism by Reference Method | NG-Test CARBA 5 Result |     |     |     |     |         |         | Grand Total |
|--|------------------------|-----|-----|-----|-----|---------|---------|-------------|
|  | KPC                    | OXA | VIM | IMP | NDM | NDM+OXA | OXA+VIM |             |
| KPC                                      | 26                     |     |     |     |     |         |         | 26          |
| OXA                                      |                        | 15  |     |     |     |         |         | 15          |
| VIM                                      |                        |     | 8   |     |     |         |         | 8           |
| IMP                                      |                        |     |     | 9   |     |         |         | 11          |
| NDM                                      |                        |     |     |     | 27  |         |         | 27          |
| NDM+OXA                                  |                        |     |     |     |     | 3       |         | 3           |
| OXA+VIM                                  |                        |     |     |     |     |         | 1       | 1           |
| Negative                                 |                        |     |     |     |     |         | 19      | 19          |
| Grand Total                              | 26                     | 15  | 8   | 9   | 27  | 3       | 1       | 110         |

Table 2. Resistance Mechanism Detection by NG-Test CARBA 5 on MAC

| Resistance Mechanism by Reference Method | NG-Test CARBA 5 Result |     |     |     |     |         |         | Grand Total |
|--|------------------------|-----|-----|-----|-----|---------|---------|-------------|
|  | KPC                    | OXA | VIM | IMP | NDM | NDM+OXA | OXA+VIM |             |
| KPC                                      | 26                     |     |     |     |     |         |         | 26          |
| OXA                                      |                        | 15  |     |     |     |         |         | 15          |
| VIM                                      |                        |     | 8   |     |     |         |         | 8           |
| IMP                                      |                        |     |     | 10  |     |         |         | 11          |
| NDM                                      |                        |     |     |     | 27  |         |         | 27          |
| NDM+OXA                                  |                        |     |     |     |     | 3       |         | 3           |
| OXA+VIM                                  |                        |     |     |     |     |         | 1       | 1           |
| Negative                                 |                        |     |     |     |     |         | 19      | 19          |
| Grand Total                              | 26                     | 15  | 8   | 10  | 27  | 3       | 1       | 110         |

Table 3. Resistance Mechanism Detection by NG-Test CARBA 5 on MH

| Resistance Mechanism by Reference Method | NG-Test CARBA 5 Result |     |     |     |     |         |         | Grand Total |
|--|------------------------|-----|-----|-----|-----|---------|---------|-------------|
|  | KPC                    | OXA | VIM | IMP | NDM | NDM+OXA | OXA+VIM |             |
| KPC                                      | 26                     |     |     |     |     |         |         | 26          |
| OXA                                      |                        | 15  |     |     |     |         |         | 15          |
| VIM                                      |                        |     | 8   |     |     |         |         | 8           |
| IMP                                      |                        |     |     | 8   |     |         |         | 11          |
| NDM                                      |                        |     |     |     | 27  |         |         | 27          |
| NDM+OXA                                  |                        |     |     |     |     | 3       |         | 3           |
| OXA+VIM                                  |                        |     |     |     |     |         | 1       | 1           |
| Negative                                 |                        |     |     |     |     |         | 19      | 19          |
| Grand Total                              | 26                     | 15  | 8   | 8   | 27  | 3       | 1       | 110         |

Figure 2. Quality Control Test Organisms




Figure 3. Isolates with >1 Mechanism Detected





Figure 4. IMP-27



## RESULTS

Table 4. Sensitivity and Specificity by Organism Type

| BAP                             | n   | True Positive | False Positive | True Negative | False Negative | Sensitivity | Specificity |
|---------------------------------|-----|---------------|----------------|---------------|----------------|-------------|-------------|
| <i>Klebsiella</i>               | 37  | 34            | 0              | 3             | 0              | 100.00%     | 100.00%     |
| <i>Enterobacter</i>             | 23  | 16            | 0              | 7             | 0              | 100.00%     | 100.00%     |
| <i>Pseudomonas</i>              | 9   | 4             | 0              | 5             | 0              | 100.00%     | 100.00%     |
| Other <i>Enterobacteriaceae</i> | 41  | 35            | 0              | 4             | 2              | 94.59%      | 100.00%     |
| Total                           | 110 | 89            | 0              | 19            | *2             | 97.80%      | 100.00%     |

| MAC                             | n   | True Positive | False Positive | True Negative | False Negative | Sensitivity | Specificity |
|---------------------------------|-----|---------------|----------------|---------------|----------------|-------------|-------------|
| <i>Klebsiella</i>               | 37  | 34            | 0              | 3             | 0              | 100.00%     | 100.00%     |
| <i>Enterobacter</i>             | 23  | 16            | 0              | 7             | 0              | 100.00%     | 100.00%     |
| <i>Pseudomonas</i>              | 9   | 4             | 0              | 5             | 0              | 100.00%     | 100.00%     |
| Other <i>Enterobacteriaceae</i> | 41  | 36            | 0              | 4             | 1              | 97.30%      | 100.00%     |
| Total                           | 110 | 90            | 0              | 19            | *1             | 98.90%      | 100.00%     |

| MH                              | n   | True Positive | False Positive | True Negative | False Negative | Sensitivity | Specificity |
|---------------------------------|-----|---------------|----------------|---------------|----------------|-------------|-------------|
| <i>Klebsiella</i>               | 37  | 34            | 0              | 3             | 0              | 100.00%     | 100.00%     |
| <i>Enterobacter</i>             | 23  | 16            | 0              | 7             | 0              | 100.00%     | 100.00%     |
| <i>Pseudomonas</i>              | 9   | 4             | 0              | 5             | 0              | 100.00%     | 100.00%     |
| Other <i>Enterobacteriaceae</i> | 41  | 34            | 0              | 4             | 3              | 91.89%      | 100.00%     |
| Total                           | 110 | 88            | 0              | 19            | #3             | 96.70%      | 100.00%     |

\* *Proteus mirabilis* (n=2) with IMP-27  
 \* *Proteus mirabilis* (n=1) with IMP-27  
 # *Proteus mirabilis* (n=2) and *Providencia rettgerii* (n=1) with IMP-27

## CONCLUSIONS

• The NG-Test CARBA 5 identified carbapenemases with 97.8%, 98.9%, and 96.7% sensitivity on BAP, MAC, and MH agar and 100% specificity on all media types.

• All resistance genes were detected with the exception of 3 strains with IMP-27 (identified by WGS) that were variably detected with the NG-Test CARBA 5 assay. IMP-27 detection varied across media; 1/3 was detected on BAP, 2/3 on MAC, and 0/3 were detected on MH agar.

• Notably, strains with >1 carbapenemase (OXA + VIM (n=1), OXA + NDM (n=3)) demonstrated 100% correlation between the NG-Test CARBA 5 and reference methods.

• The NG-Test CARBA 5 is sensitive and specific for KPC, OXA-48-like, VIM, IMP, and NDM. With a 15 minute incubation period, it can easily be performed in real time on the bench when resistance to a carbapenem is first detected and could streamline workflow.

## Acknowledgements

The supplies for this study were provided by Hardy Diagnostics.